



# CERTIFICATE OF CALIBRATION



0653

**Date of Issue: 03 January 2024**

**Certificate Number: UCRT24/1014**

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: [info@noise-and-vibration.co.uk](mailto:info@noise-and-vibration.co.uk)

Web: [www.noise-and-vibration.co.uk](http://www.noise-and-vibration.co.uk)

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages



Customer TNEI  
7th Floor West One  
Forth Banks  
Newcastle Upon Tyne  
NE1 3PA

Order No. 5001

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00643024
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	43052
	Rion	Microphone	UC-59	06804
	Rion	Calibrator	NC-75	34334830
		Calibrator adaptor type if applicable		NC-75-022

Performance Class 1

Test Procedure TP 2.SLM 61672-3 TPS-49

*Procedures from IEC 61672-3:2006 were used to perform the periodic tests.*

Type Approved to IEC 61672-1:2002 YES Approval Number 21.21 / 13.02

*If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003*

Date Received 22 December 2023

ANV Job No. UKAS23/12871

Date Calibrated 03 January 2024

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

Previous Certificate	Dated	Certificate No.	Laboratory
	23 November 2023	214284	NSAI National Metrology Lab.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# CERTIFICATE OF CALIBRATION

Certificate Number

UCRT24/1014

UKAS Accredited Calibration Laboratory No. 0653

Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title	Sound Level Meter	NL-42 / NL-52
SLM instruction manual ref / issue		11-03
SLM instruction manual source	Manufacturer	
Internet download date if applicable	N/A	
Case corrections available	Yes	
Uncertainties of case corrections	Yes	
Source of case data	Manufacturer	
Wind screen corrections available	Yes	
Uncertainties of wind screen corrections	Yes	
Source of wind screen data	Manufacturer	
Mic pressure to free field corrections	Yes	
Uncertainties of Mic to F.F. corrections	Yes	
Source of Mic to F.F. corrections	Manufacturer	
Total expanded uncertainties within the requirements of IEC 61672-1:2002	Yes	
Specified or equivalent Calibrator	Specified	
Customer or Lab Calibrator	Lab Calibrator	
Calibrator adaptor type if applicable	NC-75-022	
Calibrator cal. date	18 December 2023	
Calibrator cert. number	UCRT23/2596	
Calibrator cal cert issued by	0653	
Calibrator SPL @ STP	94.01	dB Calibration reference sound pressure level
Calibrator frequency	1000.00	Hz Calibration check frequency
Reference level range	25 - 130	dB

Accessories used or corrected for during calibration - None

Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

Environmental conditions during tests	Start	End	
Temperature	22.68	23.06	± 0.30 °C
Humidity	45.1	53.7	± 3.00 %RH
Ambient Pressure	97.79	97.77	± 0.03 kPa

Response to associated Calibrator at the environmental conditions above.

Initial indicated level	94.1	dB	Adjusted indicated level	94.0	dB
The uncertainty of the associated calibrator supplied with the sound level meter ±				0.10	dB

Self Generated Noise This test is currently not performed by this Lab.

Microphone installed (if requested by customer) = Less Than	N/A	dB	A Weighting
Uncertainty of the microphone installed self generated noise ±	N/A	dB	

Microphone replaced with electrical input device -	UR = Under Range indicated			
Weighting	A	C	Z	
	11.1	15.5	20.8	
	dB	dB	dB	
	UR	UR	UR	

Uncertainty of the electrical self generated noise ± 0.12 dB

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

END

Calibrated by:

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

None

R 1





# CERTIFICATE OF CALIBRATION



0653

**Date of Issue: 05 August 2024**

**Certificate Number: UCRT24/2059**

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: [info@noise-and-vibration.co.uk](mailto:info@noise-and-vibration.co.uk)

Web: [www.noise-and-vibration.co.uk](http://www.noise-and-vibration.co.uk)

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory

Customer TNEI  
7th Floor West One  
Forth Banks  
Newcastle Upon Tyne  
NE1 3PA

Order No.	5001																												
Description	Sound Level Meter / Pre-amp / Microphone / Associated Calibrator																												
Identification	<table><thead><tr><th>Manufacturer</th><th>Instrument</th><th>Type</th><th>Serial No. / Version</th></tr></thead><tbody><tr><td>Rion</td><td>Sound Level Meter</td><td>NL-52</td><td>00520923</td></tr><tr><td>Rion</td><td>Firmware</td><td></td><td>2.1</td></tr><tr><td>Rion</td><td>Pre Amplifier</td><td>NH-25</td><td>11770</td></tr><tr><td>Rion</td><td>Microphone</td><td>UC-59</td><td>21320</td></tr><tr><td>Rion</td><td>Calibrator</td><td>NC-75</td><td>34334830</td></tr><tr><td></td><td>Calibrator adaptor type if applicable</td><td></td><td>NC-75-022</td></tr></tbody></table>	Manufacturer	Instrument	Type	Serial No. / Version	Rion	Sound Level Meter	NL-52	00520923	Rion	Firmware		2.1	Rion	Pre Amplifier	NH-25	11770	Rion	Microphone	UC-59	21320	Rion	Calibrator	NC-75	34334830		Calibrator adaptor type if applicable		NC-75-022
Manufacturer	Instrument	Type	Serial No. / Version																										
Rion	Sound Level Meter	NL-52	00520923																										
Rion	Firmware		2.1																										
Rion	Pre Amplifier	NH-25	11770																										
Rion	Microphone	UC-59	21320																										
Rion	Calibrator	NC-75	34334830																										
	Calibrator adaptor type if applicable		NC-75-022																										

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013

*Procedures from IEC 61672-3:2013 were used to perform the periodic tests.*

Type Approved to IEC 61672-1:2013 Yes

*If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013*

Date Received 02 August 2024

ANV Job No. UKAS24/08565

Date Calibrated 05 August 2024

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate

Dated

Certificate No.

Laboratory

Initial Calibration

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

<b>CERTIFICATE OF CALIBRATION</b>	<b>Certificate Number</b>
	<b>UCRT24/2059</b>
UKAS Accredited Calibration Laboratory No. 0653	Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title	NL-52/NL-42 Description for IEC 61672-1		
SLM instruction manual ref / issue	No. 56034 21-03	Source	Rion
Date provided or internet download date	19 March 2021		
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013			YES
Specified or equivalent Calibrator	Specified		
Customer or Lab Calibrator	Lab Calibrator		
Calibrator adaptor type if applicable	NC-75-022		
Calibrator cal. date	10 July 2024		
Calibrator cert. number	UCRT24/1958		
Calibrator cal cert issued by Lab	0653		
Calibrator SPL @ STP	94.00	dB	Calibration reference sound pressure level
Calibrator frequency	1000.00	Hz	Calibration check frequency
Reference level range	Single	dB	
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			

Environmental conditions during tests		Start	End	
Temperature		24.48	24.36	± 0.30 °C
Humidity		62.2	59.0	± 3.00 %RH
Ambient Pressure		100.04	100.02	± 0.03 kPa

Indication at the Calibration Check Frequency			
Initial indicated level	94.1	dB	Adjusted indicated level 94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB

Self Generated Noise	
Microphone installed -	Less Than 20.7 dB A Weighting
Microphone replaced with electrical input device -	UR = Under Range indicated
Weighting	A C Z
	13.8 dB UR 19.6 dB UR 25.6 dB UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

None

..... END .....  
 Calibrated by: [REDACTED] R 3





## CERTIFICATE OF CALIBRATION

**Date of Issue: 22 August 2024**

**Certificate Number: TCRT24/1658**

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: [info@noise-and-vibration.co.uk](mailto:info@noise-and-vibration.co.uk)

Web: [www.noise-and-vibration.co.uk](http://www.noise-and-vibration.co.uk)

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages

Customer TNEI Services Ltd  
7th Floor  
West One  
Forth Banks  
Newcastle Upon Tyne  
NE13PA

Order No. 5001

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00721000
	Rion	Firmware		2.1
	Rion	Pre Amplifier	NH-25	22106
	Rion	Microphone	UC-59	21938
	Rion	Calibrator	NC-74	34762316
		Calibrator adaptor type if applicable		NC-74-002

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013

*Procedures from IEC 61672-3:2013 were used to perform the periodic tests.*

Type Approved to IEC 61672-1:2013 Yes

*If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013*

Date Received 21 August 2024

ANV Job No. TRAC24/08373

Date Calibrated 22 August 2024

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
Initial Calibration			

This certificate provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# CERTIFICATE OF CALIBRATION



Certificate Number

TCRT24/1658

Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title	NL-52/NL-42 Description for IEC 61672-1		
SLM instruction manual ref / issue	No. 56034 21-03	Source	Rion
Date provided or internet download date	19 March 2021		
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013			YES
Specified or equivalent Calibrator	Specified		
Customer or Lab Calibrator	Customers Calibrator		
Calibrator adaptor type if applicable	NC-74-002		
Calibrator cal. date	14 February 2024		
Calibrator cert. number	UCRT24/1237		
Calibrator cal cert issued by Lab	ANV Measurement Systems		
Calibrator SPL @ STP	94.03	dB	Calibration reference sound pressure level
Calibrator frequency	1002.41	Hz	Calibration check frequency
Reference level range	Single	dB	
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			

Environmental conditions during tests	Start	End	
Temperature	22.98	22.72	± 0.30 °C
Humidity	52.4	56.3	± 3.00 %RH
Ambient Pressure	99.47	99.46	± 0.03 kPa

Indication at the Calibration Check Frequency																			
Initial indicated level		94.3			dB		Adjusted indicated level		94.0		dB								
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±										0.10		dB							
Self Generated Noise																			
Microphone installed -		Less Than		19.7		dB		A Weighting											
Microphone replaced with electrical input device -							UR = Under Range indicated												
Weighting		A				C				Z									
		10.4		dB		UR		15.8		dB		UR		22.8		dB		UR	

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with the Guide to the Expression of Uncertainty in Measurement published by ISO.

## Additional Comments

None

END

Calibrated by:

R 2


## Appendix E – Noise Modelling Data

---


[REDACTED]

[REDACTED]





[illegible]





[REDACTED]

[REDACTED]








[REDACTED]


[REDACTED]



## Appendix F – Figures

---





LEGEND

Site Boundary

Area Excluded from Site Boundary

Noise Monitoring Locations (NMLs)

NML01

NML02

NML03

NML04

Nearest Identified Noise Sensitive Receptors (NSRs)

NSRs Represented by NML01

NSRs Represented by NML02

NSRs Represented by NML03

NSRs Represented by NML04

NSR Not Included Within Assessment

0	13/02/2025	FOR PLANNING	KB	EW
Rev.	Date	Amendment Details	Drawn	Approved

Contains OS data © Crown Copyright and database right 2022

tnei

FIELD

Client

FOR INFORMATION

Project Title:

FYRISH BESS

Drawing Title:

FIGURE 01 - NOISE STUDY AREA

Scale:

1:8,000

Original Size:

A3


Spatial Reference:

British National Grid

Drawing Number:

16819-004





LEGEND

Site Boundary

Area Excluded from Site Boundary

Noise Assessment Locations (NALs)

Modelled Noise Sources

Modelled Buildings

Modelled Acoustic Fence (5 m)

Modelled Acoustic Fence (4 m)

Predicted Noise Levels (dBA)

25 - 30

30 - 35

35 - 40

40 - 45

45 - 50

50 - 55

55 - 60

60 - 65

Noise contours modelled in accordance with ISO 9613 Part 2:2024 at a height of 4 m and displayed on a 10 m by 10 m grid.

All noise sources assumed to be operating concurrently.

All levels shown as dB LAeq(t).

0	12/02/2025	FOR PLANNING	KB	EW
Rev.	Date	Amendment Details	Drawn	Approved

Contains OS data © Crown Copyright and database right 2022

tnei

FIELD

Client

Drawing Status: FOR INFORMATION

Project Title: FYRISH BESS

Drawing Title: FIGURE 02 - NOISE CONTOUR PLOT

Scale: 1:6,500

Original Size: A3

Spatial Reference: British National Grid

Drawing Number: 16819-005

## Appendix G – One-Third Octave Band Predicted Levels (dBZ)

Noise Assessment Location (NAL)	External Predicted Noise Levels, dB(Z)																											
	25	31.5	40	50	63	80	100	125	160	200	250	315	400	500	630	800	1000	1250	1600	2000	2500	3150	4000	5000	6300	8000	10000	
NAL01 – Culcraggie Lodge	37	34	33	36	31	28	26	26	29	31	24	28	29	28	26	25	-	-	-	-	-	-	-	-	-	-	-	-
NAL02 – Clashnabuiac	40	36	35	37	32	29	27	26	28	30	24	28	27	26	24	22	20	18	16	-	-	-	-	-	-	-	-	-
NAL03 – Deer Park Cottages	34	30	29	31	27	24	19	18	20	23	17	21	21	20	18	17	15	13	-	-	-	-	-	-	-	-	-	-
NAL04 – Dairy Bungalow	35	31	30	32	28	25	21	20	22	25	18	22	23	22	20	19	17	15	13	10	6	-	-	-	-	-	-	-
NAL05 – Fyrish	32	29	28	30	26	23	19	19	21	24	17	21	21	21	19	17	15	13	11	7	2	-	-	-	-	-	-	-
Where a dash (-) is presented, predicted values were negligibly low (0 dB or below), and as such were not included within the table.																												